

## Results and conclusion<sup>2</sup>

In accordance with the **EN 13624 (2021)** and the "Requirements and methods for **VAH**-certification of chemical disinfection processes" (**2015/2022**), the batch B23001 of the test product **Sporosan® WD 42**, when applied at the concentration / contact time - relation of at least **50 % / 15 s** under clean conditions (0.03 % albumin), **possesses yeasticidal efficacy** ( $\log_{10} \text{RF} \geq 5$ ) for the reference strain *C. albicans* (Tab. 1.2).

In accordance with the **EN 13624 (2021)** and the "Requirements and methods for **VAH**-certification of chemical disinfection processes" (**2015/2022**), the batch B23001 of the test product **Sporosan® WD 42**, when applied at the concentration / contact time - relation of at least **80 % / 15 s** under clean conditions (0.03 % albumin), **possesses yeasticidal and fungicidal efficacy** ( $\log_{10} \text{RF} \geq 5$ ) for the reference strains *C. albicans* and *A. brasiliensis* (Tab. 1.2 - 2).

Results are considered validated in accordance with the requirements of the **EN 13624 (2021)** and **VAH (2015/2022)**.

Greifswald, August 31, 2023

  
Dr. rer med. (Dipl. Biol.) T. Koburger-Banssen  
- General Manager -

  
Prof. Dr. med. A. Kramer  
for Hygiene and Environmental Medicine -

  
KREDITIERTES PROFLAB  
VAH  
HYGIENE NORD GMBH  
W.-RATHENAU-STRASSE  
17489 GREIFSWALD  
NACH DIN EN ISO/IEC 17025