

Results and conclusions ²

In accordance with the **EN 14348 (2005)** and the "Requirements and methods for **VAH**-certification of chemical disinfection processes" (**2015/2023**), the test product **MEIKO®ACTIVE SPOROSAN® 8100**, when applied at a concentration / contact time relation of at least **97 % / 30 s** at **20±1 °C** under clean conditions (0.03 % albumin), **possesses tuberculocidal efficacy** ($\log_{10} \text{RF} \geq 4$) for the reference strain *M. terrae* (table 1.1 – 1.2).

In accordance with the **EN 14348 (2005)** and the "Requirements and methods for **VAH**-certification of chemical disinfection processes" (**2015/2023**), the test product **MEIKO®ACTIVE SPOROSAN® 8100**, when applied at a concentration / contact time relation of at least **97 % / 30 s** at **20±1 °C** under clean conditions (0.03 % albumin), **possesses mycobactericidal efficacy** ($\log_{10} \text{RF} \geq 4$) for the reference strains *M. terrae* and *M. avium* (table 1.1 – 2.2).

Greifswald, October 16, 2024


Dr. rer. med. (Dipl. Biol.) T. Koburger-Janssen
- General Manager -


Prof. Dr. med. A. Kramer
MD for Hygiene and Environmental Medicine -