

Results and conclusion ²

In accordance with the **EN 17387 (2021)**, the batches H 023005/2 and N 023005/1 of the 2-component-disinfectant **MEIKO®ACTIVE SPOROSAN® 8100**, when applied as described above at the concentration / contact time - relation of at least **50 % / 1 min** at **20 ± 1 °C** under clean conditions (0.03 % albumin), **possesses tuberculocidal efficacy** ($\log_{10} \text{RF} \geq 4$) for reference strain *M. terrae* (Tab. 1).


In accordance with the **EN 17387 (2021)**, the batches H 023005/2 and N 023005/1 of the 2-component-disinfectant **MEIKO®ACTIVE SPOROSAN® 8100**, when applied as described above at the concentration / contact time - relation of at least **50 % / 1 min** at **20 ± 1 °C** under clean conditions (0.03 % albumin), **possesses mycobactericidal efficacy** ($\log_{10} \text{RF} \geq 4$) for reference strain *M. terrae* and *M. avium* (Tab. 1).

Results are considered validated in accordance with the requirements of **EN 17387 (2021)**.

Greifswald, February 03, 2025


Dr. rer. med. (Dipl. Biol.) T. Koburger-Janssen
- General Manager -


SENE NORD GMBH
W.-RATHENAU-STR. 49A
17489 GREIFSWALD
NACH DIN EN ISO/IEC 17025


Prof. Dr. med. A. Kramer
MD for Hygiene and Environmental Medicine -