


Results and conclusion²

In accordance with the **EN 13624 (2021)** and the “Requirements and methods for **VAH**-certification of chemical disinfection processes” (**2015/2025**), the batches B25002/ B25001 of the 2-component-product **MEIKO® ACTIVE SPOROSAN® 8100 A+B**, when applied at the concentration / contact time - relation of at least **80 % / 15 s** under clean conditions (0.03 % albumin), **possesses yeasticidal efficacy** ($\log_{10} \text{RF} \geq 4$) for the reference strain *C. albicans* (Tab. 1).

In accordance with the **EN 13624 (2021)** and the “Requirements and methods for **VAH**-certification of chemical disinfection processes” (**2015/2025**), the batches B25002/ B25001 of the 2-component-product **MEIKO® ACTIVE SPOROSAN® 8100 A+B**, when applied at the concentration / contact time - relation of at least **80 % / 15 s** under clean conditions (0.03 % albumin), **possesses fungicidal efficacy** ($\log_{10} \text{RF} \geq 4$) for the reference strains *C. albicans* and *A. brasiliensis* (Tab. 1 - 2).

Results are considered validated in accordance with the requirements of the **DIN EN 13624 (2021)** and **VAH (2015/2025)**.

Greifswald, March 30, 2026


Dr. rer. med. (Dipl. Biol.) T. Koburger-Janssen
- General Manager -




Prof. Dr. med. A. Kramer
- MD for Hygiene and Environmental Medicine -